

Kim Reynolds, Governor

Kelly Garcia, Director

April 12, 2025

Daniel Roach 7402 COUNTY ROAD E SOUTH RANGE, WI 54874

Dear Daniel Roach,

This letter is to verify that you have met all MQSA initial qualification requirements as stated in the final regulations, 900.12(a)(3)(i) and all Iowa registration requirements for a medical physicist in: Mammography: Digital Mammography: Tomosynthesis

Therefore, you are permitted to perform all those procedures required under Iowa Radiation Machines Rules for the above categories. Your **registration number MPHY10101** expires on April 30, 2026.

Each Iowa facility where you provide medical physics services must have a copy of this Medical Physics Approval letter.

Thank you for your cooperation. Please call 641-227-2585 if you have any questions.

Sincerely,

Patty Riesberg

Patty Riesberg, Bureau Chief Bureau of Radiological Health Office Phone: 515-371-2255 Email: patty.riesberg@hhs.iowa.gov



November 7, 2024 ATTESTATION REGARDING INITIAL REQUIREMENTS OF THE MAMMOGRAPHY QUALITY STANDARDS ACT AND/OR ACR REQUIREMENTS FOR DIGITAL AND DBT BREAST IMAGING

This document is intended to provide proof of medical physicist's initial qualification in Digital and Tomosynthesis (DBT) Mammography.

Attestation must include as much of the following information as possible:

Name of the institution/facility where the applicable training or mammography reading/interpreting, or other activity, took place; name of the course(s) or training (where applicable); the attendance, reading/interpreting, or other activity dates; and the supervising/responsible person (where applicable) for the institution/facility.

I, Steven Nicholas, attest that, to the best of my knowledge and my belief, the following information provided in this declaration is true and correct. Under my direct supervision, Daniel Roach, MS, a Radiation Physics Consultants, Inc. physicist, has met the Initial Qualifications requirements of MQSA and the FDA with 3 hours of Digital Mammography training and 28 hours of DBT Mammography training.

"Have a master's degree or higher in a physical science with at least 20 semester hours (30 quarter hours) of graduate or undergraduate physics, and, have the experience of conducting surveys of at least one mammography facility with a total of at least 10 mammography units, and at least 20 hours of mammography facility survey training."

Please see the additional details on the following pages.

Please do not hesitate to contact me if you have any additional questions.

Sincerely,

Steven T. Nicholas, M.S., DABMP President, RPC



Facility	Type of Unit	Description of Tests	Time (hrs)	Date
Essentia Health Hospital (Graceville, MN)	DBT	Annual Physics Survey	3	1/19/23
Bigfork Valley Hospital (Bigfork, MN)	DBT	Annual Physics Survey	3.00	2/23/23
Essentia Health Clinic (Grand Rapids, MN)	DBT	Annual Physics Survey	3.5	3/3/23
Essentia Health Hospital (Moose Lake, MN)	DBT	Annual Physics Survey	3.00	12/20/23
CHI St. Joe's Hospital (Park Rapids, MN)	DBT	Annual Physics Survey	3.00	12/22/23
Essentia Health Clinic (Hermantown, MN)	DBT	Annual Physics Survey	3.00	1/5/24
Essentia Health Clinic (Detroit Lakes, MN)	DBT	Annual Physics Survey	3.50	1/25/24
Essentia Health (Sandstone, MN)	DBT	Annual Physics Survey	3.50	2/14/24
Essentia Health Clinic (l'Falls, MN)	DBT	Annual Physics Survey	2.50	7/25/24
CMDI (Pine City, MN)	Digital	Annual Physics Survey	3	7/29/24

Total DBT (hrs):	28
Total Digital (hrs):	3



Site Name Bigfork Valley Hospital				Rep	ort Date	e 11/7/2024				
Address		258 Pine Tree Drive, Bigfork, MN			Surv	vey Date	(23, 11		
Medical P	hysicist's Name	Steve Nicholas and Danny Roach (Training)			Si	gnature	E E	Hen TV alos		
X-Ray Uni	it Manufacturer	Lorad/Holo	gic			Model	Sele	Selenia Dimensions DBT		
Date of In	stallation	5/16/201	8		F	Room ID		Mammog	aphy	
						SN		81002132	2146	
QC Manua	al Version #	MAN-03706, Rev. 010	(August	2020)	(use any ve	rsion applica	able to mode	l; contact mfi	r if questions)	
Accessor	y Equipment	Manufacturer	Мо	odel	Loca	ation	Q	C Manual V	ersion #	
	Review Workstation*	Barco/Hologic	MDM	G-5221	√ On-site	Off-site	M	AN-02568, I	Rev. 002	
	Film Printer*	NA	١	١A	N	A		NA		
FDA's Policy	Guidance Help Syste	ors and printers specifically clear m (www.accessdata.fda.gov/cc	lrh_docs/pi	resentations	/pghs/Polic	_Guidance	e_Help_Sys	stem.htm).	_	
Survey Ty Features	/pe ⊡	Mammo Eqpt Evaluation (MEE 2D	,	nit (include l omosynthesi	-	its for Man	nmo Eqpt c	hecklist) 🗋	Annual Survey	
 Collin Artifa kVp A Beam Evalu Autor Breas Aver 	nation Assessmen ct Evaluation Accuracy and Rep o Quality Assessm ation of System R natic Exposure Co of Entrance Expos rage glandular dose	roducibility ent - HVL Measurement	/ and Av mGy (300	e rage Gla) mrad) (co	ndular D	ose	113 130	mrad mrad	PASS/FAIL Pass Pass Pass Pass Pass Pass Pass Pas	
9. Radia	tion Output Rate								Pass	
10. Phant	tom Image Quality	/ Evaluation	Fibers	Specks	Masses					
ANN	Phantom image s	cores (conventional)	6.0	4.0	4.5				Pass	
	Phantom image s	scores (DBT)	5.5	4.0	4.0				Pass	
11. Signa	I-To-Noise Ratio a	and Contrast-To-Noise R	atio Mea	surement	t s (values i	required fo	r all tests)			
	SNR (value)	51.9							Pass	
	CNR (value)	10.76 (required for I			ual Survey))				
		ary by more than ±15% (۸/							Pass	
-		rkstation (RWS) QC (for al	I RWS, eve	en if located	offsite; NA	if only har	dcopy read	0	Pass	
	M Printer QC (if app	• •							NA	
	4. Detector Flat Field Calibration (MEE only)						NA			
		or Tomosynthsis (DBT M	EE only)	C					NA	
-		s Indicator (MEE only)							Pass	
	pression (MEE only)								NA	
18. Detec	8. Detector Ghosting (troubleshooting only)								NA	

*** YOUR MEDICAL PHYSICIST MUST SUMMARIZE HIS/HER RESULTS ON THIS FORM ***

(Lorad, continued)

Evaluation of Site's Technologist QC Program

		Frequency	PASS/FAIL
1.	DICOM Printer Quality Control (if applicable)	Weekly	NA
2.	Viewboxes and Viewing Conditions	Weekly	Pass
3.	Artifact Evaluation	Weekly	Pass
4.	Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	Pass
5.	Phantom Image Quality Evaluation	Weekly	Pass
6.	Detector Flat-Field Calibration	Weekly	Pass
7.	Compression Thickness Indicator	Bi-weekly	Pass
8.	Visual Checklist	Monthly	Pass
9.	Repeat/Reject Analysis	Quarterly	Pass
10.	Compression	Semi-annually	Pass
11.	. Geometry Calibration (Tomosynthsis Option) (DBT)	Semi-annually	Pass
12	Diagnostic Review Workstation QC (NA if only hardcopy read)	See Hologic QC Manual	Pass
13.	Mobile Unit Quality Control (if applicable)	After every move	NA

Medical Physicist's Recommendations for Quality Improvement

This is a Medical Physicist's annual survey. There was also a software upgrade performed on 2/11/2022. It did not require an MEE but let this report serve as oversight and verification.

Medical Physicist's QC Tests

No Discrepancies.

Evaluation of Site's Technologist QC Program

There are no discrepancies.

Site does not print.

Site Name	Name Essentia Health - St. Mary's Clinic				2/20/2024
Address	1027 Was	hington Ave., Detroit Lakes	, MN 56501	Survey Date	1/25/2024
Medical Physicist's Name		Shane McCotter & Danny Roach (training)		Signature	Sh-Mbob
X-Ray Unit	Manufacturer	Lorad/Ho	logic	Model	Selenia Dimensions
Date of Installation		3/27/20	15	Room ID	Mammagraphy Room 1005
				SN	81002154467
QC Manual	Version #	MAN-03706 Rev. 0	11 (Nov. 2021)	(use any version applic	cable to model; contact mfr if questions)
Accessory	Equipment	Manufacturer	Model	Location	QC Manual Version #
F	Review Workstation*	Barco/Hologic	MDMC-12133	On-site	MAN-03706 Rev. 011 (Nov. 2021)
	Film Printer*	NA	NA	NA	NA

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polic_Guidance_Help_System.htm).

Survey Type: Annual Survey

Features: 2D & Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

PASS/FAIL

1.	I. Mammographic Unit Assembly Evaluation						
2.							
3.	Artifact Evaluation	Pass					
4.	kVp Accuracy and Reproducibility	Pass					
5.	Beam Quality Assessment - HVL Measurement	Pass					
6.	Evaluation of System Resolution	Pass					
7.	Automatic Exposure Control (AEC) Function Performance (NA for systems without AEC)	Pass					
8.	Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose						
AN	Average glandular dose for average breast is ≤3 mGy (300 mrad) <i>(conventional)</i> 125 mrad	Pass					
Ś	, Average glandular dose for average breast is ≤3 mGy (300 mrad) <i>(DBT)</i> 153 mrad	Pass					
9.	Radiation Output Rate	Pass					
10.	Phantom Image Quality Evaluation Fibers Specks Masses						
	Phantom image scores <i>(conventional)</i> 6.0 4.0 4.5	Pass					
	Phantom image scores (DBT) 6.0 4.0 4.0	Pass					
11.	Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements (values required for all tests)						
	SNR (value) 56.6	Pass					
	CNR (value) 11.52 (required for new unit MEE and Annual Survey)						
	CNR should not vary by more than ±15% (NA for MEE)	Pass					
12.	12. Diagnostic Review Workstation (RWS) QC (for all RWS, even if located offsite; NA if only hardcopy read)						
13.	3. DICOM Printer QC (if applicable, MEE only)						
14.	4. Detector Flat Field Calibration (MEE only)						
15.	5. Geometry Calibration For Tomosynthsis (DBT MEE only)						
16.	Compression Thickness Indicator (MEE only)	NA					
17.	7. Compression (MEE only)						
18.	8. Detector Ghosting (troubleshooting only)						

(Lorad, continued)

Evaluation of Site's Technologist QC Program

(Required for Annual Surveys; not required for Mammography Equipment Evaluations of new units. However, medical physicists must review the site's technologist QC program within 45 days and complete this section so that the facility may submit this form along with the entire Mammography Equipment Evaluation report with their phantom and clinical images to the ACR.)

		Frequency	PASS/FAIL
1.	DICOM Printer Quality Control (if applicable)	Weekly	NA
2.	Viewboxes and Viewing Conditions	Weekly	Pass
3.	Artifact Evaluation	Weekly	Pass
4.	Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	Pass
5.	Phantom Image Quality Evaluation	Weekly	Pass
6.	Detector Flat-Field Calibration	Weekly	Pass
7.	Compression Thickness Indicator	Bi-weekly	Pass
8.	Visual Checklist	Monthly	Pass
9.	Repeat/Reject Analysis	Quarterly	Pass
10). Compression	Semi-annually	Pass
11	. Geometry Calibration (Tomosynthsis Option) (DBT, 🖉	Semi-annually	Fail
12	2. Diagnostic Review Workstation QC (NA if only hardcopy read)	See Hologic QC Manual	Pass
13	B. Mobile Unit Quality Control (if applicable)	After every move	NA

Medical Physicist's Recommendations for Quality Improvement

This is a Medical Physicist's annual survey.

Medical Physicist's QC Tests

No Discrepancies.

Evaluation of Site's Technologist QC Program

We observed the Geometry Calibration QC was not run within 6 months in 2023. Was completed on 11/1/2022 then again on 7/5/2023. Did not perform again in 2023.

Site Name	Ess	sentia Health Sandstone Hos	Report Date	3/14/2024	
Address	705 Lu	undorff Drive, Sandstone, MN	55072	Survey Date	2/14/2024
Medical Phy	/sicist's Name	Steven Nicholas & Danny Roach (Training)		Signature	Sten Thinks
X-Ray Unit M	Manufacturer	Lorad/Hold	ogic	Model	Selenia 3Dimensions DBT
Date of Installation		1/6/2022		Room ID	Mammography
	-			SN	3DM160101808
QC Manual	Version #	MAN-03706, Rev. 0	11 (Nov 2021)	(use any version applied	cable to model; contact mfr if questions)
Accessory E	Equipment	Manufacturer	Model	Location	QC Manual Version #
R	eview Workstation*	Barco	MDNG-13221	Offsite	MAN-03706, Rev. 011 (Nov 2021)
	Film Printer*	NA	NA	NA	NA

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polic_Guidance_Help_System.htm).

> Survey Type: Annual Survey

Features: 2D & Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

1.	Mammographic Unit Assembly Evaluation						Pass
2.	Collimation Assessment						Pass
3.	Artifact Evaluation						Pass
4.	kVp Accuracy and Reproducibility						Pass
5.	Beam Quality Assessment - HVL Measuremen	t					Pass
6.	Evaluation of System Resolution						Pass
7.	Automatic Exposure Control (AEC) Function P	Performa	nce (NA fo	or systems	without A	EC)	Pass
8.	Breast Entrance Exposure, AEC Reproducibili	ty and A	verage G	landular	Dose		
M	、Average glandular dose for average breast is ≤3	mGy (30	00 mrad) ((conventior	nal)	126 mrad	Pass
Ś	I Average glandular dose for average breast is ≤3	6 mGy (30	0 mrad)	(DBT)		143 mrad	Pass
9.	Radiation Output Rate						Pass
10.	Phantom Image Quality Evaluation	Fibers	Specks	Masses			
	Phantom image scores (conventional)	6.0	4.0	4.5			Pass
	Phantom image scores (DBT)	6.0	4.0	4.5			Pass
11.	Signal-To-Noise Ratio and Contrast-To-Noise	Ratio Me	easureme	ents (value	es require	ed for all tests)	
	SNR (value) 55.3						Pass
	CNR (value) 10.78 (required for	new unit N	IEE and An	nual Surve	y)		
	CNR should not vary by more than ±15%	(NA for MI	EE)				Pass
12.	2. Diagnostic Review Workstation (RWS) QC (for all RWS, even if located offsite; NA if only hardcopy read)						Pass
13.	3. DICOM Printer QC (if applicable, MEE only)						NA
14.	4. Detector Flat Field Calibration (MEE only)						NA
15.	5. Geometry Calibration For Tomosynthsis (DBT MEE only)						NA
16.	6. Compression Thickness Indicator (MEE only)						NA
17.	17. Compression (MEE only)						NA
18.	18. Detector Ghosting (troubleshooting only)						NA

*** YOUR MEDICAL PHYSICIST MUST SUMMARIZE HIS/HER RESULTS ON THIS FORM ***

PASS/FAIL

(Lorad, continued)

Evaluation of Site's Technologist QC Program

(Required for Annual Surveys; not required for Mammography Equipment Evaluations of new units. However, medical physicists must review the site's technologist QC program within 45 days and complete this section so that the facility may submit this form along with the entire Mammography Equipment Evaluation report with their phantom and clinical images to the ACR.)

		Frequency	PASS/FAIL
1.	DICOM Printer Quality Control (if applicable)	Weekly	Pass
2.	Viewboxes and Viewing Conditions	Weekly	Pass
3.	Artifact Evaluation	Weekly	Pass
4.	Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	Pass
5.	Phantom Image Quality Evaluation	Weekly	Pass
6.	Detector Flat-Field Calibration	Weekly	Pass
7.	Compression Thickness Indicator	Bi-weekly	Pass
8.	Visual Checklist	Monthly	Pass
9.	Repeat/Reject Analysis	Quarterly	Pass
10	. Compression	Semi-annually	Pass
11	. Geometry Calibration (Tomosynthsis Option) (DBT, 🖉	Semi-annually	Pass
12	. Diagnostic Review Workstation QC (NA if only hardcopy read)	See Hologic QC Manual	Pass
13	. Mobile Unit Quality Control (if applicable)	After every move	Pass

Medical Physicist's Recommendations for Quality Improvement

This is a Medical Physicist's annual survey.

Medical Physicist's QC Tests

No Discrepancies.

Evaluation of Site's Technologist QC Program

No Discrepancies.

Site Name Es	ssentia Health- Grand Rapids C	Clinic		Repo	ort Date		4/2/20	023	
Address 1542 G	Golf Course Rd, Grand Rapids, MN 55744			Surv	ey Date	() 1 3/3/2023, 1 A			
Medical Physicist's Name	Steven T. Nicholas and Danny Roach (trainee)			Sig	nature	S	Lilo la		
X-Ray Unit Manufacturer	Lorad/Hold	ogic			Model	Sel	enia Dime	nsions DBT	
Date of Installation	3/2/2018	8		R	oom ID	Mar	nmo Room	n (Suite 204)	
					SN		SDM1315		
QC Manual Version #	MAN-03706, Rev. 00	09 (Sept 2	2019)	(use any ver	sion applicabl	le to mode	el; contact m	nfr if questions)	
Accessory Equipment	Manufacturer	M	odel	Locat	ion	Ç	C Manual	Version #	
Review Workstation*	Hologic	Sec	urView	🗆 On-site 🛛	Off-site	IAM	N-03706, Rev. (009 (Sept 2019)	
Film Printer*	NA	1	NA	NA	\ \		NA	N .	
*FDA recommends that only monito Policy Guidance Help System (www Survey Type [□]		s/presentat	ions/pghs/P	olic_Guidano					
Features 🛛 🖉	2D 🛛 Digita	al Breast To	omosynthesi	s (DBT)					
 Evaluation of System R Automatic Exposure Co Breast Entrance Expos Average glandular dose 	oducibility ent - HVL Measurement	erforman y and Ave mGy (300	e rage Gla) mrad) <i>(c</i> o	ndular Do		122	mrad	Pass Pass Pass Pass Pass Pass Pass Pass	
9. Radiation Output Rate			,	,		104		Pass	
10. Phantom Image Quality	v Evaluation	Fibers	Specks	Masses					
Phantom image s	COres (conventional)	6.0	4.0	4.5				Pass	
Phantom image s	cores (DBT)	6.0	4.0	4.5				Pass	
11. Signal-To-Noise Ratio a	and Contrast-To-Noise R	atio Mea	surement	s (values re	equired for a	ll tests)			
SNR (value)	57.6							Pass	
CNR (value)	11.18 (required for	new unit M	EE and Ann	ual Survey)					
CNR should not v	rary by more than ±15% <i>(I</i>	VA for MEE)					Pass	
 Diagnostic Review Wor 	kstation (RWS) QC (for a	ll RWS, eve	en if located	offsite; NA if	only hardco	py read)		Pass	
13. DICOM Printer QC (if app	• • •							NA	
14. Detector Flat Field Calibration (MEE only)							NA		
•	•	EE only)	C					Pass	
16. Compression Thicknes	15. Geometry Calibration For Tomosynthsis (DBT MEE only)								
·- • ·								Pass	
17. Compression (MEE only) 18. Detector Ghosting (trouk								NA NA	

(Lorad, continued)

Evaluation of Site's Technologist QC Program

		Frequency	PASS/FAIL
1.	DICOM Printer Quality Control (if applicable)	Weekly	NA
2.	Viewboxes and Viewing Conditions	Weekly	Pass
	Artifact Evaluation	Weekly	Pass
4.	Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	Pass
5.	Phantom Image Quality Evaluation	Weekly	Pass
6.	Detector Flat-Field Calibration	Weekly	Pass
7.	Compression Thickness Indicator	Bi-weekly	Pass
8.	Visual Checklist	Monthly	Pass
9.	Repeat/Reject Analysis	Quarterly	Pass
10.	Compression	Semi-annually	Pass
11.	Geometry Calibration (Tomosynthsis Option) (DBT)	Semi-annually	Pass
	Diagnostic Review Workstation QC (NA if only hardcopy read)	See Hologic QC Manual	Pass
13.	Mobile Unit Quality Control (if applicable)	After every move	NA

Medical Physicist's Recommendations for Quality Improvement

This is annual testing.		
Medical Physicist's QC Tests		
No Discrepancies.		
Evaluation of Site's Technologist QC Program		
No discrepancies.		
Facility does not print hard copy.		

Site Name	Esse	ntia Health - Holy Trinity H	ospital	Report Date	2/15/2023			
Address	115 W	/est 2nd St, Graceville, MN	1 56240	Survey Date				
Medical Physicist's Name		Shane McCotter & Danny Roach (Training)		Signature	Sh-Mitoto			
X-Ray Unit	Manufacturer	Lorad/Ho	logic	Model	Selenia Dimensions			
Date of Installation		9/12/2018		Room ID	Mammography Room #1913			
				SN	SDM131900425			
QC Manual	Version #	MAN-03706 Rev. 007 (March 2018)		(use any version appli	cable to model; contact mfr if questions)			
Accessory	Equipment	Manufacturer	Model	Location	QC Manual Version #			
F	Review Workstation* Barco/Hologic SecureView		SecureView	Off-site	MAN-03706 Rev. 007 (March 2018)			
Film Printer* NA NA		NA	NA	NA				
*EDA recomm	EDA recommends that only monitors and printers specifically cleared for EEDM use by EDA's Office of Device Evaluation (ODE) be used. See							

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polic_Guidance_Help_System.htm).

Survey Type: Annual Survey

Features: 2D & Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

PASS/FAIL

Mammographic Unit Assembly Evaluation Pass 1. **Collimation Assessment** 2. Pass 3. Artifact Evaluation Pass 4. kVp Accuracy and Reproducibility Pass 5. Beam Quality Assessment - HVL Measurement Pass 6. Evaluation of System Resolution Pass 7. Automatic Exposure Control (AEC) Function Performance (NA for systems without AEC) Pass 8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose Average glandular dose for average breast is ≤3 mGy (300 mrad) (conventional) Pass 131 mrad Average glandular dose for average breast is ≤3 mGy (300 mrad) (*DBT*) 161 mrad Pass 9. Radiation Output Rate Pass 10. Phantom Image Quality Evaluation Fibers Specks Masses Phantom image scores (conventional) 5.5 4.0 4.5 Pass Phantom image scores (DBT) 4.0 5.0 4.5 Pass 11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements (values required for all tests) SNR (value) 57.0 Pass CNR (value) 11.06 (required for new unit MEE and Annual Survey) CNR should not vary by more than ±15% (NA for MEE) Pass 12. Diagnostic Review Workstation (RWS) QC (for all RWS, even if located offsite; NA if only hardcopy read) Pass **13. DICOM Printer QC** (*if applicable, MEE only*) NA **14. Detector Flat Field Calibration** (MEE only) NA 15. Geometry Calibration For Tomosynthsis (DBT MEE only) NA 16. Compression Thickness Indicator (MEE only) Pass **17.** Compression (MEE only) Pass **18. Detector Ghosting** (troubleshooting only) NA

(Lorad, continued)

Evaluation of Site's Technologist QC Program

(Required for Annual Surveys; not required for Mammography Equipment Evaluations of new units. However, medical physicists must review the site's technologist QC program within 45 days and complete this section so that the facility may submit this form along with the entire Mammography Equipment Evaluation report with their phantom and clinical images to the ACR.)

		Frequency	PASS/FAIL
1.	DICOM Printer Quality Control (if applicable)	Weekly	NA
2.	Viewboxes and Viewing Conditions	Weekly	Pass
3.	Artifact Evaluation	Weekly	Pass
4.	Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	Pass
5.	Phantom Image Quality Evaluation	Weekly	Pass
6.	Detector Flat-Field Calibration	Weekly	Pass
7.	Compression Thickness Indicator	Bi-weekly	Pass
8.	Visual Checklist	Monthly	Pass
9.	Repeat/Reject Analysis	Quarterly	Pass
10	. Compression	Semi-annually	Pass
11	. Geometry Calibration (Tomosynthsis Option) (Semi-annually	Pass
12	. Diagnostic Review Workstation QC (NA if only hardcopy read)	See Hologic QC Manual	Pass
13	. Mobile Unit Quality Control (if applicable)	After every move	NA

Medical Physicist's Recommendations for Quality Improvement

This is a Medical Physicist's annual survey.

Medical Physicist's QC Tests

No Discrepancies. Images are read at Essentia Health - Fargo Hospital. RWS testing performed annually.

Evaluation of Site's Technologist QC Program

No Discrepancies.

Site Name	Esse	entia Health - Hermantown	Report Date	1/30/2024	
Address	4855W. Ai	4855W. Arrowhead Rd, Hermantown, MN 55811 Survey Da		Survey Date	1/5/2024
Medical Physicist's Name		Shane McCotter & Danny Roach (training)		Signature	Sh-Mfot
X-Ray Unit Manufacturer		Lorad/Hologic		Model	3Dimensions
Date of Installation		12/2/2021		Room ID	Mammo Room
				SN	3DM160101598
QC Manual Version #		MAN-03706 Rev. 011 (Nov. 2021)		(use any version appli	cable to model; contact mfr if questions)
Accessory	Equipment	Manufacturer	Model	Location QC Manual Version #	
Review Workstation*		Barco	MDMC-12133	Off-site	MAN-03706 Rev. 010 (Qug 2020)
Film Printer*		NA	NA	NA	NA

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polic_Guidance_Help_System.htm).

> Survey Type: Annual Survey

Features: 2D & Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

- 1. Mammographic Unit Assembly Evaluation 2. Collimation Assessment 3. Artifact Evaluation 4. kVp Accuracy and Reproducibility 5. Beam Quality Assessment - HVL Measurement 6. Evaluation of System Resolution 7. Automatic Exposure Control (AEC) Function Performance (NA for systems without AEC) 8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose Average glandular dose for average breast is ≤3 mGy (300 mrad) (conventional) Pass 123 mrad Average glandular dose for average breast is ≤3 mGy (300 mrad) (*DBT*) 152 mrad Pass 9. Radiation Output Rate Pass **10.** Phantom Image Quality Evaluation Fibers Specks Masses Phantom image scores (conventional) 6.0 4.0 4.5 Pass Phantom image scores (DBT) 4.0 6.0 4.5 Pass 11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements (values required for all tests) SNR (value) Pass 55.2 CNR (value) 10.98 (required for new unit MEE and Annual Survey) CNR should not vary by more than ±15% (NA for MEE) 12. Diagnostic Review Workstation (RWS) QC (for all RWS, even if located offsite; NA if only hardcopy read) **13. DICOM Printer QC** (*if applicable, MEE only*) **14. Detector Flat Field Calibration** (MEE only) 15. Geometry Calibration For Tomosynthsis (DBT MEE only) **16.** Compression Thickness Indicator (MEE only) 17. Compression (MEE only)
- **18. Detector Ghosting** (troubleshooting only)

*** YOUR MEDICAL PHYSICIST MUST SUMMARIZE HIS/HER RESULTS ON THIS FORM ***

PASS/FAIL
Pass

Pass
Pass
NA

(Lorad, continued)

Evaluation of Site's Technologist QC Program

(Required for Annual Surveys; not required for Mammography Equipment Evaluations of new units. However, medical physicists must review the site's technologist QC program within 45 days and complete this section so that the facility may submit this form along with the entire Mammography Equipment Evaluation report with their phantom and clinical images to the ACR.)

Frequency	PASS/FAIL
Weekly	NA
Weekly	NA
Weekly	Pass
Bi-weekly	Pass
Monthly	Pass
Quarterly	Pass
Semi-annually	Pass
Semi-annually	Pass
) See Hologic QC Manual	Pass
After every move	NA
	Weekly Weekly Weekly Weekly Weekly Bi-weekly Bi-weekly Monthly Quarterly Semi-annually Semi-annually

Medical Physicist's Recommendations for Quality Improvement

This is a Medical Physicist's annual survey.

Medical Physicist's QC Tests

No Discrepancies.

Evaluation of Site's Technologist QC Program

No Discrepancies.

Site Name	Essen	tia Health International Falls	s Clinic	Report Date	8/8/2024	
Address	2501 Kee	nan Dr, International Falls,	MN 56649	Survey Date	7/25/2024	
Medical Physicist's Name		Shane McCotter & Danny Roach (training)		Signature	Sh-Mfot	
X-Ray Unit Manufacturer		Lorad/Hologic		Model	Selenia	
Date of Installation		1/16/2015		Room ID	Mammo Room	
				SN	2841214W8278W	
QC Manual Version #		MAN-01476 Rev. 002 Sept 2014		(use any version applicable to model; contact mfr if question		
Accessory	Equipment	Manufacturer	Model	Location	QC Manual Version #	
F	Review Workstation*	Off-Site	Hologic	SecurView	MAN-01476 Rev. 001	
Film Printer*		NA NA		NA	NA	
	•	s and printers specifically cl (www.accessdata.fda.gov/c		-	e Evaluation (ODE) be used. See e_Help_System.htm).	

Survey Type: Annual Survey Features: 2D

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

PASS/FAIL 1. Mammographic Unit Assembly Evaluation Pass **Collimation Assessment** 2. Pass 3. Artifact Evaluation & Detector Uniformity Pass 4. kVp Accuracy and Reproducibility Pass 5. Beam Quality Assessment - HVL Measurement Pass 6. Evaluation of System Resolution Pass 7. Automatic Exposure Control (AEC) Function Performance (NA for systems without AEC) Pass 8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose Average glandular dose for average breast is <3 mGy (300 mrad) (conventional) mrad Pass 128 Average glandular dose for average breast is ≤3 mGy (300 mrad) (*DBT*) NA mrad NA 9. Radiation Output Rate Pass 10. Phantom Image Quality Evaluation Fibers Specks Masses Phantom image scores (conventional) 5.0 4.0 4.5 Pass Phantom image scores (DBT) NA NA NA NA 11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements (values required for all tests) SNR (value) Pass 56.6 CNR (value) 11.36 (required for new unit MEE and Annual Survey) CNR should not vary by more than ±15% (NA for MEE) Pass 12. Diagnostic Review Workstation (RWS) QC (for all RWS, even if located offsite; NA if only hardcopy read) Pass **13. DICOM Printer QC** (if applicable, MEE only) NA **14. Detector Flat Field Calibration** (MEE only) NA 15. Geometry Calibration For Tomosynthesis (DBT MEE only) NA **16.** Compression Thickness Indicator (MEE only) NA **17.** Compression (MEE only) NA **18. Detector Ghosting** (troubleshooting only) NA 19. Upright Biopsy Phantom Image Quality Evaluation NA 20. Upright Biopsy QAS Evaluation NA

(Lorad, continued)

Evaluation of Site's Technologist QC Program

(Required for Annual Surveys; not required for Mammography Equipment Evaluations of new units. However, medical physicists must review the site's technologist QC program within 45 days and complete this section so that the facility may submit this form along with the entire Mammography Equipment Evaluation report with their phantom and clinical images to the ACR.)

Frequency	PASS/FAIL
Weekly	NA
Weekly	Pass
Bi-weekly	Pass
Monthly	Pass
Quarterly	Pass
Semi-annually	Pass
Semi-annually	Pass
See Hologic QC Manual	Pass
After every move	NA
	Weekly Weekly Weekly Weekly Weekly Bi-weekly Monthly Quarterly Semi-annually Semi-annually

Medical Physicist's Recommendations for Quality Improvement

This is a Medical Physicist's annual survey.

Medical Physicist's QC Tests

No Discrepancies.

Evaluation of Site's Technologist QC Program

No Discrepancies.

Site Name	Esse	entia Health Moose Lake H	Report Date	1/2/2024	
Address	4572 0	4572 Co. Rd. 61, Moose Lake, MN 55767			12/20/2023
Medical Physicist's Name		Shane McCotter & Danny Roach (training)		Signature	Sh-Mft
X-Ray Unit Manufacturer		Lorad/Hologic		Model	Selenia Dimenions DBT
Date of Installation		1/20/2017		Room ID	Mammo Room
				SN	81012167628
QC Manual Version #		MAN-03706 Rev. 011 (Nov. 2021)		(use any version applica	able to model; contact mfr if questions)
Accessory	Equipment	Manufacturer	Model	Location	QC Manual Version #
Review Workstation*		Barco MDMC-12133		Off-site	MAN-04426 Rev. 001
Film Printer*		NA NA		NA	NA
* = = = = = = = = = = = = = = = = = = =	ende thet ender meetiter				- Evaluation (ODE) has wood Soo

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polic_Guidance_Help_System.htm).

Survey Type: Annual Survey

Features: 2D & Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

PASS/FAIL

Mammographic Unit Assembly Evaluation Pass 1. 2. Collimation Assessment Pass 3. Artifact Evaluation Pass 4. kVp Accuracy and Reproducibility Pass 5. Beam Quality Assessment - HVL Measurement Pass 6. Evaluation of System Resolution Pass 7. Automatic Exposure Control (AEC) Function Performance (NA for systems without AEC) Pass 8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose Average glandular dose for average breast is ≤3 mGy (300 mrad) (conventional) Pass 122 mrad Average glandular dose for average breast is ≤3 mGy (300 mrad) (*DBT*) 149 mrad Pass 9. Radiation Output Rate Pass 10. Phantom Image Quality Evaluation Fibers Specks Masses Phantom image scores (conventional) 6.0 4.0 4.5 Pass Phantom image scores (DBT) 4.0 5.5 4.0 Pass 11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements (values required for all tests) SNR (value) 57.2 Pass CNR (value) 12.89 (required for new unit MEE and Annual Survey) CNR should not vary by more than ±15% (NA for MEE) Pass 12. Diagnostic Review Workstation (RWS) QC (for all RWS, even if located offsite; NA if only hardcopy read) Pass **13. DICOM Printer QC** (*if applicable, MEE only*) NA **14. Detector Flat Field Calibration** (MEE only) NA **15.** Geometry Calibration For Tomosynthsis (DBT MEE only) NA 16. Compression Thickness Indicator (MEE only) NA **17.** Compression (MEE only) NA **18. Detector Ghosting** (troubleshooting only) NA

(Lorad, continued)

Evaluation of Site's Technologist QC Program

(Required for Annual Surveys; not required for Mammography Equipment Evaluations of new units. However, medical physicists must review the site's technologist QC program within 45 days and complete this section so that the facility may submit this form along with the entire Mammography Equipment Evaluation report with their phantom and clinical images to the ACR.)

Frequency	PASS/FAIL
Weekly	NA
Weekly	Pass
Bi-weekly	Pass
Monthly	Pass
Quarterly	Pass
Semi-annually	Pass
Semi-annually	Pass
See Hologic QC Manual	Pass
After every move	NA
	Weekly Weekly Weekly Weekly Weekly Bi-weekly Monthly Quarterly Semi-annually Semi-annually

Medical Physicist's Recommendations for Quality Improvement

This is a Medical Physicist's annual survey.

Medical Physicist's QC Tests

No Discrepancies.

Evaluation of Site's Technologist QC Program

No Discrepancies.

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY Full-Field Digital – Siemens

		i dii i	licia Bigi				
Site Name	CHI - S	t. Joseph's Health	Hospital	Repo	ort Date	1/2	/2024
Address	600 Pleasant Ave., Park Rapids, MN 56470 Survey Date		12/2	2/2023			
Medical P	hysicist's Name	Shane McCotter a	nd Danny Roach	Sic	nature	M"	ML
	t Manufacturer	Sier	nens		Model	Mammom	at Revelation
Date of In		-	/2018	R	oom ID		Vammo
			2010	Serial N			219
QC Manua	al Version #	Tomo QC 56.01.2	24, 2D QC 51.01.24	(use version applicable t	L		-
Accessory	y Equipment	Manufacturer	Model	Location		QC Manual Ver	sion #
	Review Workstation*	Barco/McKesson	5421 HD	On-Site	MA	N-01476 Rev. 001	(June 2009)
	Film Printer*	NA	NA	NA		NA	
				FFDM use by FDA's O s/presentations/pghs/P			
Survey Ty	-						
Features	2D & Digi	tal Breast Tomosy	nthesis (DBT)				
		Madi	cal Physic	ist's QC Tes	ete		
("Pass" mea	ans all components of			omponent fails. Tests m		ne for both on and	off-site equipment.)
							PASS/FAIL
1. Image	Quality						Pass
	•		and 4 masses v				
				ole if spatial resolution a	i -	,	
	Phantom image	scores:	Fibers 5.0	Specks 4.0	Masses	4.0	
	ct Detection						Pass
	and CNR Measure	ments					Pass
-	at/Reject Analysis						Pass
	ression Force	1-)					Pass
							NA
7. SNR,	CNR and AEC Rep Measured value		59.85 CNR	2.75			Pass
		l entrance air ke		2.75			Pass
			—	thin ±15% of mean	for more	uromonte	Pass
8. Radia	tion Dose	n mean pixei vai	ues and Sink wi	Init ±15% of mean	IUI IIIeas	urements	Pass
o. Naula		llar dose for ave	rane breast is <3	mGy (300 mrad)	Г	0.750 mGy	1 033
9 Snatia	al Resolution		uge biedst is <u><</u> 0		L	0.750 1109	Pass
10. AEC							Pass
	ctor Uniformity						Pass
	nanical Tests						Pass
	sition Workstatio	n Monitor Chec	k				Pass
-	udit/Evaluation of						Pass
	nation, Dead Spac	•	-	tion			Fail
	nd Radiation Out	-					Pass
	Voltage Measuren		bility				Pass
	ge Glandular Dos	-	-				Pass
19. Geom	etric Accuracy in	X and Y Directi	on and Z-Resol	ution (DBT) (Option	al Revelati	ion)	Pass
20. Radia	tion Field (DBT)						Fail
21. Syste	m Imaging Quality	(DBT)					Pass
	≥ 4 fibers, ≥ 3 s	peck groups and	≥ 3 ma <u>sses mu</u>	st be visible			
	Phantom image	scores:	Fibers 5.0	Specks 4.0	Masses	4.0	
	ct Detection (DBT)						Pass
22 Povio	w Workstation (P)	NS) Tosts (for of	I PINS over if loop	tod offaito: NIA if anly h	ordoony ro	ad	Pace

23. Review Workstation (RWS) Tests (for all RWS, even if located offsite; NA if only hardcopy read)

Pass

(Siemens, continued)

Evaluation of Technologist QC Program

New units: Medical physicists *must* review the technologist QC *within 45 days of installation* and complete this section. The facility is required to submit the entire Mammography Equipment Evaluation report (including this form) along with their testing materials for accreditation.

Existing units: Medical physicists must complete this section as part of the unit's annual survey.

Relocating units: This section is **not** required if the medical physicist does **not** conduct a complete annual survey after relocation.

		FREQUENCY	PASS/FAIL
1.	Phantom Image Quality Novation	on & Fusion-Daily; Inspiration-Weekly	Pass
2.	Artifact Detection	Weekly	Pass
3.	SNR and CNR Measurements	Weekly	Pass
4.	Detector Calibration*	Novation-Weekly; Fusion-Quarterly	Pass
5.	Repeat/Reject Analysis	Quarterly	Pass
6.	Compression Force	Semi-annually	Pass
7.	System Imaging Quality (DBT)	Weekly	Pass
8.	Printer Check (if applicable)	Daily, when images printed	NA
9.	Review Workstation QC-Overall (NA if only hardcopy rea	d) See FDA guidance	Pass
10	. Mobile Unit Quality Control (if applicable)	After every move	NA
* Fo	r Mammomat Revelation and Inspiration indicate NA-calibration require	d before ΩC but does not need to be documented	

* For Mammomat Revelation and Inspiration, indicate NA-calibration required before QC but does not need to be documented

Medical Physicist's Recommendations for Quality Improvement

This is a Medical Physicist's annual survey.

Medical Physicist's QC Tests

Item 15: The X-Ray Field to Light Field accuracy failed to be within the required 2% of the SID (must not exceed 13 mm combined). We measured the Left/Right combined deviation to be 14mm mainly from the right side where the light field measured 10 mm greater than the x-ray field. Please have a Service Engineer review the results on Page 12 and make the necessary corrections within 30 days of the testing.

Item 15: The chest Wall Missing Tissue measured 5.1mm which is greater than the 5mm allowed limit set by Siemens. Please have a Service Engineer review and make the necessary adjustments, possibly tightening the cover to decrease the missing tissue.

Item 20: We could easily see the edge of the tomo paddle in the first projection image. Per the Siemens manual, the edge of the collimator should not be seen in any projection image. Please have a Service Engineer make the necessary adjustments.

Evaluation of Site's Technologist QC Program

No Discrepancies.

Site Name	CMDI at Welia Health Clinic			Report Date	8/15/2024	
Address	1425	N Main St, Pine City, MN	55063	Survey Date	7/29/2024	
Medical Physicist's Name		Shane McCotter & Danny Roach (training)		Signature	Sh-M/A	
X-Ray Unit Manufacturer		Lorad/Hologic		Model	Selenia Dimensions	
Date of Installation		7/10/2019		Room ID	Mammo	
				SN	SDM131900771	
QC Manual Version # MAN-03706, Rev. 011 (Nov. 2021)			(use any version applic	cable to model; contact mfr if questions)		
Accessory I	Equipment	Manufacturer	Model	Location	QC Manual Version #	
R	eview Workstation*	Barco/Hologic	MDMC-12133	Off-Site	MAN-04959, Rev. 002	
	Film Printer*	NA	NA	NA	NA	
*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polic_Guidance_Help_System.htm).						

Survey Type: Annual Survey

Features: 2D & Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

PASS/FAIL 1. Mammographic Unit Assembly Evaluation Pass **Collimation Assessment** 2. Pass 3. Artifact Evaluation & Detector Uniformity Pass 4. kVp Accuracy and Reproducibility Pass 5. Beam Quality Assessment - HVL Measurement Pass 6. Evaluation of System Resolution Pass 7. Automatic Exposure Control (AEC) Function Performance (NA for systems without AEC) Pass 8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose Average glandular dose for average breast is <3 mGy (300 mrad) (conventional) mrad Pass Average glandular dose for average breast is ≤3 mGy (300 mrad) (*DBT*) 121 146 mrad Pass 9. Radiation Output Rate Pass 10. Phantom Image Quality Evaluation Fibers Specks Masses Phantom image scores (conventional) 5.0 4.0 4.0 Pass Phantom image scores (DBT) 5.0 4.0 4.0 Pass 11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements (values required for all tests) SNR (value) Pass 56.8 CNR (value) 11.38 (required for new unit MEE and Annual Survey) CNR should not vary by more than ±15% (NA for MEE) Pass 12. Diagnostic Review Workstation (RWS) QC (for all RWS, even if located offsite; NA if only hardcopy read) Pass **13. DICOM Printer QC** (if applicable, MEE only) NA **14. Detector Flat Field Calibration** (MEE only) NA 15. Geometry Calibration For Tomosynthesis (DBT MEE only) NA **16.** Compression Thickness Indicator (MEE only) NA **17.** Compression (MEE only) NA **18. Detector Ghosting** (troubleshooting only) NA 19. Upright Biopsy Phantom Image Quality Evaluation NA 20. Upright Biopsy QAS Evaluation NA

(Lorad, continued)

Evaluation of Site's Technologist QC Program

(Required for Annual Surveys; not required for Mammography Equipment Evaluations of new units. However, medical physicists must review the site's technologist QC program within 45 days and complete this section so that the facility may submit this form along with the entire Mammography Equipment Evaluation report with their phantom and clinical images to the ACR.)

Frequency	PASS/FAIL
Weekly	NA
Weekly	Pass
Bi-weekly	Pass
Monthly	Pass
Quarterly	Pass
Semi-annually	Pass
Semi-annually	Pass
See Hologic QC Manual	Pass
After every move	NA
	Weekly Weekly Weekly Weekly Weekly Bi-weekly Monthly Quarterly Semi-annually Semi-annually

Medical Physicist's Recommendations for Quality Improvement

This is a Medical Physicist's annual survey.

Medical Physicist's QC Tests

No Discrepancies.

Evaluation of Site's Technologist QC Program

No Discrepancies.

John Patrick Aniversity Mealth and Applied Sciences

Apon recommendation of the Faculty,

John Patrick University of Health and Applied Sciences has conferred upon

DANIEL ROACH

the degree of

MASTER OF SCIENCE IN MEDICAL PHYSICS

Who has honorably fulfilled all the requirements prescribed by the University for that degree

at South Bend, Indiana this twenty-first day of August in the year of our Lord two thousand and twenty-three



John Patrick University of Health and Applied Sciences

Official Transcript

100 E. Wayne Street, Suite 140, South Bend, IN 46601 Phone: (574)232-2408, Fax: (574)232-2200

RECIPIENT:

Daniel Roach 1101 N 57th Ave. W Duluth, MN 55907

Degrees/Certificates

Master of Science in Medical Physics Granted 8/21/2023 Transcrint

STUDENT: Roach, Daniel Student ID: 2022000214 Birthdate: May 22, 1998 Enrollment Date: Sep 6, 2021

MP502 Physics of Radiation Biology 3.00 3.00 B 9.00 MP590 Medical and Professional Ethics 1.00 1.00 A 4.0 Totals 8.00 8.00 Term GPA: 3.63 Cum. GPA: 3.6 2021-2022: Spring 2022 - 01/10/2022 - 04/25/2022 Attempted Cr. Grade Point MP503 Physics of Radiation Oncology I 3.00 3.00 B 9.0 MP505 Physics of Radiation Oncology I 3.00 3.00 B 9.0 MP503 Physics of Radiation Oncology I 3.00 3.00 B 9.0 MP505 Physics of Radiation Oncology I 3.00 3.00 A 4.0 Totals Totals 7.00 4.00 Term GPA: 3.25 Cum. GPA: 3.5 2021-2022: Summer 2022 - 05/09/2022 - 08/22/2022 Curse # Name S.00 2.00 A 8.0 MHP601 Shielding Design 2.00 2.00 A 8.0 8.0 D222-2023: Fall 2022 - 09/05/2022 - 12/19/2022 Grade Poin	Transcri					
BIOL530 Human Anatomy & Physiology 4.00 4.00 A.00 A.00 A.00 M.00			Attempted Cr	Farnad Cr	Crada	Doints
MP590 Medical and Professional Ethics 1.00 1.00 A 4.0 Totals 8.00 8.00 Term GPA: 3.63 Cum. GPA: 3.63 2021-2022: Spring 2022 - 01/10/2022 - 04/25/2022 Attempted Cr. Earned Cr. Grade Point MP503 Physics of Diagnostic Radiology 3.00 3.00 B 9.00 MP505 Physics of Radiation Oncology I 3.00 3.00 B 9.00 MP599 S1 Seminars Session 1 1.00 1.00 A.00 Term GPA: 3.25 Cum. GPA: 3.63 2021-2022: Surmer 2022 - 05/09/2022 - 08/22/2022 Course # Name Attempted Cr. Earned Cr. Grade Point MIP601 Shielding Design 2.00 2.00 A 8.00 MP503 Physics of Diagnostic Radiology 3.00 3.00 A 12.00 Totals Sum Attempted Cr. Earned Cr. Grade Point MIP601 Shielding Design 3.00 3.00 A 12.00 Dotof Stal Name <td></td> <td></td> <td>1</td> <td></td> <td></td> <td>16.00</td>			1			16.00
Totals 8.00 Term GPA: 3.63 Cum. GPA: 3.63 2021-2022: Spring 2022 - 01/10/2022 - 04/25/2022 Attempted Cr. Earned Cr. Grade Point MP503 Physics of Diagnostic Radiology 3.00 - W - MP505 Physics of Radiation Oncology I 3.00 3.00 B 9.00 MP599 S1 Semiaras Session 1 1.00 1.00 A.00 Ferm GPA: 3.25 Cum. GPA: 3.05 2021-2022: Summer 2022 - 05/09/2022 - 08/22/2022 7.00 4.00 Term GPA: 3.25 Cum. GPA: 3.05 2021-2022: Summer 2022 - 05/09/2022 - 08/22/2022 Cume GPA: 3.00 A.00 A.00 Ferm GPA: 3.05 Cum. GPA: 3.63 2021-2022: Summer 2022 - 05/09/2022 - 08/22/2022 Cume GPA: 3.00 A.00	MP502	Physics of Radiation Biology	3.00	3.00	В	9.00
2021-2022: Spring 2022 - 01/10/2022 - 04/25/2022 Attempted Cr. Earned Cr. Grade Point MP503 Physics of Diagnostic Radiology 3.00 W W	MP590	Medical and Professional Ethics	1.00	1.00	А	4.00
Course # MP503 Name Physics of Diagnostic Radiology Attempted Cr. 3.00 Earned Cr. 0.00 Grade W Point W MP505 Physics of Radiation Oncology I 3.00 3.00 3.00 3.00 9.00 MP505 Physics of Radiation Oncology I 3.00 3.00 3.00 A 9.00 MP509 S1 Seminars Session 1 1.00 1.00 1.00 A 4.00 Totals	Totals		8.00	8.00	Term GPA: 3.63	Cum. GPA: 3.63
MP503 Physics of Diagnostic Radiology 3.00 - W MP505 Physics of Radiation Oncology I 3.00 3.00 B 9.00 MP505 Seminars Session 1 1.00 1.00 A. 4.00 Totals 7.00 4.00 Term GPA: 3.25 Cum. GPA: 3.5 2021-2022: Summer 2022 - 05/09/2022 - 08/22/2022 - - - - Course # Name Attempted Cr. Earned Cr. Grade Point MIP503 Physics of Diagnostic Radiology 3.00 3.00 A 12.00 Totals 5.00 5.00 Term GPA: 4.00 Cum. GPA: 3.6 2022-2023: Full 2022 - 09/05/2022 - 12/19/2022 - - - - Course # Name Attempted Cr. Earned Cr. Grade Point MHP510 Hadiation Oncology II 3.00 3.00 A 12.00 MP506 Physics of Radiation Oncology II 3.00 3.00 A 12.00 MP513 Physics of Nuclear On	2021-2022: 8	Spring 2022 - 01/10/2022 - 04/25/2022				
MP505 Physics of Radiation Oncology I 3.00 3.00 B 9.00 MP505 Seminars Session 1 1.00 1.00 A 4.00 Totals 7.00 4.00 Term GPA: 3.25 Cum. GPA: 3.5 2021-2022: Summer 2022 - 05/09/2022 - 08/22/2022 - Earned Cr. Grade Point MHP601 Shielding Design 2.00 2.00 A 8.00 MP503 Physics of Diagnostic Radiology 3.00 3.00 A 12.00 Totals 5.00 5.00 Ferm GPA: 4.00 Cum. GPA: 3.60 2022-2023: Fall 2022 - 09/05/2022 - 12/19/2022 - - - - Course # Name Attempted Cr. Grade Point MHP510 Health Physics and Radiation Safety 3.00 3.00 A 12.00 MP506 Physics of Nuclear Oncology II 3.00 3.00 A 12.00 Totals 9.00 9.00 Term GPA: 4.00 Cum. GPA: 3.7 2022-2023: Spring 2023 - 01/09/2023 - 04/24/2023				Earned Cr.		Points
MP599 S1 Seminars Session 1 1.00 1.00 1.00 4.00 4.00 Totals 7.00 4.00 Term GPA: 3.25 Cum. GPA: 3.5 2021-2022: Summer 2022 - 05/09/2022 - 08/22/2022 Attempted Cr. Earned Cr. Grade Point Ourse # Name Attempted Operation 2.00 2.00 A 8.0 MP503 Physics of Diagnostic Radiology 3.00 3.00 3.00 A 12.0 Totals 5.00 5.00 Term GPA: 4.00 Cum. GPA: 3.6 2022-2023: Fall 2022 - 09/05/2022 - 12/19/2022 Course # Name Attempted Cr. Earned Cr. Grade Point MHP510 Health Physics and Radiation Safety 3.00 3.00 A 12.0 MP506 Physics of Nuclear Oncology I 3.00 3.00 A 12.0 MP613 Physics of Nuclear Oncology I 3.00 3.00 A 12.0 MP504 Physics of Nuclear Medicine 3.00 3.00 A 12.0 MP508 <	MP503	Physics of Diagnostic Radiology	3.00		W	
Totals 7.00 4.00 Term GPA: 3.25 Cum. GPA: 3.5 2021-2022: Summer 2022 - 05/09/2022 - 08/22/2022 Attempted Cr. Earned Cr. Grade Point MHP601 Shielding Design 2.00 2.00 A 8.0 MP503 Physics of Diagnostic Radiology 3.00 3.00 A 12.0 Totals 5.00 5.00 Ferm GPA: 4.00 Cum. GPA: 3.6 2022-2023: Fall 2022 - 09/05/2022 - 12/19/2022 Cum. GPA: 3.00 3.00 A 12.0 Course # Name Attempted Cr. Earned Cr. Grade Point MHP510 Health Physics and Radiation Safety 3.00 3.00 A 12.0 MP506 Physics of Nuclear Oncology II 3.00 3.00 A 12.0 MP613 Physics of Nuclear Oncology I 3.00 3.00 A 12.0 Totals Point 3.00 3.00 A 12.0 MP614 Physics of Nuclear Medicine 3.00 3.00 A 12.0	MP505	Physics of Radiation Oncology I	3.00	3.00	В	9.00
2021-2022: Summer 2022 - 05/09/2022 - 08/22/2022 Course # Name Attempted Cr. Earned Cr. Grade Point MHP601 Shielding Design 2.00 3.00 3.00 A 12.00 MP503 Physics of Diagnostic Radiology 3.00 3.00 A 12.00 Totals 5.00 5.00 Term GPA: 4.00 Cum. GPA: 3.60 2022-2023: Fall 2022 - 09/05/2022 - 12/19/2022 Course # Name Attempted Cr. Earned Cr. Grade Point MHP510 Health Physics and Radiation Safety 3.00 3.00 A 12.00 MP506 Physics of Radiation Oncology II 3.00 3.00 A 12.00 MP613 Physics of Nuclear Oncology 3.00 3.00 A 12.00 Course # Name MP504 Physics of Nuclear Medicine 3.00 3.00 A 12.00 MP508 Radiological Instrumentation 2.00 2.00 A 8.00 MP508 Radio	MP599 S1	Seminars Session 1	1.00	1.00	A	4.00
Course #NameAttempted Cr.Earned Cr.GradePointMHP601Shielding Design2.002.00A8.0MP503Physics of Diagnostic Radiology3.003.00A12.0Totals5.005.00Term GPA: 4.00Cum. GPA: 3.62022-2023: Fall 2022 - 09/05/2022 - 12/19/2022Course #NameAttempted Cr.Earned Cr.GradePointMHP510Health Physics and Radiation Safety3.003.00A12.0MP506Physics of Radiation Oncology II3.003.00A12.0MP613Physics of Nuclear Oncology3.003.00A12.0Totals9.009.00Term GPA: 4.00Cum. GPA: 3.72022-2023: Spring 2023 - 01/09/2023 - 04/24/2023Course #NameAttempted Cr.Earned Cr.GradeMP508Radiological Instrumentation2.002.00A8.0MP509Steminars Session 101.001.00A4.0MP603Advanced Diagnostic Radiology2.002.00B6.0	Totals		7.00	4.00	Term GPA: 3.25	Cum. GPA: 3.50
MHP601 Shielding Design 2.00 2.00 A 8.0 MP503 Physics of Diagnostic Radiology 3.00 3.00 3.00 A 12.0 Totals 5.00 5.00 Term GPA: 4.00 Cum. GPA: 3.6 2022-2023: Fall 2022 - 09/05/2022 - 12/19/2022 - - Grade Point MHP510 Health Physics and Radiation Safety 3.00 3.00 A 12.0 MP506 Physics of Radiation Oncology II 3.00 3.00 A 12.0 MP613 Physics of Nuclear Oncology 3.00 3.00 A 12.0 Totals - 9.00 9.00 Term GPA: 4.00 Cum. GPA: 3.7 2022-2023: Spring 2023 - 01/09/2023 - 04/24/2023 - - 9.00 9.00 Term GPA: 4.00 Cum. GPA: 3.7 2022-2023: Spring 2023 - 01/09/2023 - 04/24/2023 - - 6.0 12.0 MP504 Physics of Nuclear Medicine 3.00 3.00 A 12.0 MP508 Radiological Instrumentation 2.00	2021-2022: 8	Summer 2022 - 05/09/2022 - 08/22/2022	101-			
MP 503 Physics of Diagnostic Radiology 3.00 3.00 3.00 A 12.0 Totals 5.00 5.00 Term GPA: 4.00 Cum. GPA: 3.6 2022-2023: Fall 2022 - 09/05/2022 - 12/19/2022 Attempted Cr. Earned Cr. Grade Point MHP510 Health Physics and Radiation Safety 3.00 3.00 A 12.0 MP506 Physics of Radiation Oncology II 3.00 3.00 A 12.0 MP613 Physics of Nuclear Oncology 3.00 3.00 A 12.0 2022-2023: Spring 2023 - 01/09/2023 - 04/24/2023 Attempted Cr. Earned Cr. Grade Point MP508 Radiological Instrumentation 2.00 3.00 3.00 A 12.0 MP508 Radiological Instrumentation 2.00 2.00 A 3.00 3.00 A 3.00 MP508 Radiological Instrumentation 2.00 2.00 A 4.00 4.00 MP509 Seminars Session 10 1.00 1.00 A 4.00						Points
Totals5.005.00Term GPA: 4.00Cum. GPA: 3.602022-2023: Fall 2022 - 09/05/2022 - 12/19/2022Course #NameAttempted Cr.Earned Cr.GradePointMHP510Health Physics and Radiation Safety3.003.00A12.00MP506Physics of Radiation Oncology II3.003.00A12.00MP613Physics of Nuclear Oncology3.003.00A12.00Totals9.009.00Term GPA: 4.00Cum. GPA: 3.702022-2023: Spring 2023 - 01/09/2023 - 04/24/2023Attempted Cr.Earned Cr.GradePointMP504Physics of Nuclear Medicine3.003.00A12.00MP508Radiological Instrumentation2.002.00A8.00MP509 S10Seminars Session 101.001.00A4.00MP603Advanced Diagnostic Radiology2.002.00B6.00		0 0				
2022-2023: Fall 2022 - 09/05/2022 - 12/19/2022Course #NameAttempted Cr.Earned Cr.GradePointMHP510Health Physics and Radiation Safety3.003.00A12.0MP506Physics of Radiation Oncology II3.003.00A12.0MP613Physics of Nuclear Oncology3.003.00A12.0Totals9.009.009.00Term GPA: 4.00Cum. GPA: 3.72022-2023: Spring 2023 - 01/09/2023 - 04/24/2023Course #NameAttempted Cr.Earned Cr.GradePointMP504Physics of Nuclear Medicine3.003.00A12.0MP508Radiological Instrumentation2.002.00A8.0MP599 S10Seminars Session 101.001.00A4.0MP603Advanced Diagnostic Radiology2.002.002.00B6.0	MP503	Physics of Diagnostic Radiology		3.00	200	12.00
Course # MHP510Name Health Physics and Radiation SafetyAttempted Cr. 3.00Earned Cr. 3.00GradePoint 12.0MP506Physics of Radiation Oncology II3.003.00A12.0MP613Physics of Nuclear Oncology3.003.00A12.0MP613Physics of Nuclear Oncology3.003.00A12.0Totals9.009.00Term GPA: 4.00Cum. GPA: 3.7Course # Physics of Nuclear MedicineAttempted Cr. 3.00GradeGradePoint 12.0MP508Radiological Instrumentation2.002.00A8.0MP599 S10Seminars Session 101.001.00A4.0MP603Advanced Diagnostic Radiology2.002.00B6.0	Totals		5.00	5.00	Term GPA: 4.00	Cum. GPA: 3.65
MHP510 Health Physics and Radiation Safety 3.00 3.00 A 12.0 MP506 Physics of Radiation Oncology II 3.00 3.00 A 12.0 MP613 Physics of Nuclear Oncology 3.00 3.00 A 12.0 MP613 Physics of Nuclear Oncology 3.00 3.00 A 12.0 Totals 9.00 9.00 Term GPA: 4.00 Cum. GPA: 3.7 2022-2023: Spring 2023 - 01/09/2023 - 04/24/2023 Attempted Cr. Grade Point MP504 Physics of Nuclear Medicine 3.00 3.00 A 12.0 MP508 Radiological Instrumentation 2.00 2.00 A 8.0 MP599 S10 Seminars Session 10 1.00 1.00 A 4.0 MP603 Advanced Diagnostic Radiology 2.00 2.00 B 6.0		Fall 2022 - 09/05/2022 - 12/19/2022				
MP506Physics of Radiation Oncology II3.003.00A12.0MP613Physics of Nuclear Oncology3.003.00A12.0Totals9.009.00Term GPA: 4.00Cum. GPA: 3.72022-2023 - 01/09/2023 - 04/24/2023Course #NameAttempted Cr.Earned Cr.GradePointMP504Physics of Nuclear Medicine3.003.00A12.0MP508Radiological Instrumentation2.002.00A8.0MP599 S10Seminars Session 101.001.00A4.0MP603Advanced Diagnostic Radiology2.002.002.00B6.0						Points
MP613Physics of Nuclear Oncology3.003.003.00A12.0Totals9.009.00Term GPA: 4.00Cum. GPA: 3.72022-2023 - 01/09/2023 - 04/24/2023Course #NameAttempted Cr.Earned Cr.GradePointMP504Physics of Nuclear Medicine3.003.00A12.0MP508Radiological Instrumentation2.002.00A8.0MP599 S10Seminars Session 101.001.001.00A4.0MP603Advanced Diagnostic Radiology2.002.00B6.0						
Totals9.009.00Term GPA: 4.00Cum. GPA: 3.72022-2023: Spring 2023 - 01/09/2023 - 04/24/2023Course #NameAttempted Cr.Earned Cr.GradePointMP504Physics of Nuclear Medicine3.003.00A12.0MP508Radiological Instrumentation2.002.00A8.0MP599 S10Seminars Session 101.001.00A4.0MP603Advanced Diagnostic Radiology2.002.00B6.0	MP506	Physics of Radiation Oncology II			A	12.00
Z022-2023: Spring 2023 - 01/09/2023 - 04/24/2023 Course # MP504 Name Physics of Nuclear Medicine Attempted Cr. 3.00 Earned Cr. 3.00 Grade A Point 12.0 MP508 Radiological Instrumentation 2.00 2.00 A 8.0 MP599 S10 Seminars Session 10 1.00 1.00 A 4.0 MP603 Advanced Diagnostic Radiology 2.00 2.00 B 6.0	MP613	Physics of Nuclear Oncology	3.00	3.00	A	12.00
Course # MP504Name Physics of Nuclear MedicineAttempted Cr. 3.00Earned Cr. 3.00GradePoint 12.0MP508Radiological Instrumentation2.002.00A8.0MP599 S10Seminars Session 101.001.00A4.0MP603Advanced Diagnostic Radiology2.002.00B6.0	Totals		9.00	9.00	Term GPA: 4.00	Cum. GPA: 3.77
MP504Physics of Nuclear Medicine3.003.00A12.0MP508Radiological Instrumentation2.002.00A8.0MP599 S10Seminars Session 101.001.00A4.0MP603Advanced Diagnostic Radiology2.002.00B6.0	2022-2023: 8	Spring 2023 - 01/09/2023 - 04/24/2023				
MP508Radiological Instrumentation2.002.00A8.0MP599 S10Seminars Session 101.001.00A4.0MP603Advanced Diagnostic Radiology2.002.00B6.0						Points
MP599 S10Seminars Session 101.001.00A4.0MP603Advanced Diagnostic Radiology2.002.00B6.0						
MP603 Advanced Diagnostic Radiology 2.00 B 6.0	MP508	Radiological Instrumentation	2.00	2.00	А	8.00
	MP599 S10	Seminars Session 10	1.00	1.00	А	4.00
Totals 8.00 8.00 Term GPA: 3.75 Cum. GPA: 3.7	MP603	Advanced Diagnostic Radiology	2.00	2.00	В	6.00
	Totals		8.00	8.00	Term GPA: 3.75	Cum. GPA: 3.76

lem LUL IV

Elizabeth M Datema Office of the Registrar


Brent D. Murphy, MS, DABR

President

2022-2023: Summer 2023 - 05/08/2023 - 08/21/2023						
Course #	Name	Attempted Cr.	Earned Cr.	Grade	Points	
MP501	Physics of Radiation Dosimetry	4.00	4.00	А	16.00	
MP699	Clinical Internship	4.00	4.00	Р	16.00	
STAT501	Statistical Methods	3.00	3.00	Α	12.00	
Totals		11.00	11.00	Term GPA: 4.00	Cum. GPA: 3.82	

Cumulative

Attempted Credits	Earned Credits	Points	GPA
48.00	45.00	172.00	3.82
0.00	0.00	0.00	0.00
48.00	45.00	172.00	3.82
	48.00 0.00	48.00 45.00 0.00 0.00	48.00 45.00 172.00 0.00 0.00 0.00



fem beth

Elizabeth M Datema Office of the Registrar

Brut Murphy Brent D. Murphy, MS, DABR

President

KEY TO TRANSCRIPT OF ACADEMIC RECORDS

Note: The following explanation reflects information found on the John Patrick University of Health and Applied Sciences (JPU) **Official Transcript** produced from the Student Information System implemented June 2011. Prior to August 5, 2019, JPU was doing business as Radiological Technologies University VT.

I. Grade and Credit Point System

The following grades are considered in computing semester or cumulative grade averages. Course hours with a grade of "F" are counted when computing grade point averages but do not count toward the earned hours required for degrees.

Graduate Courses		Undergraduate Courses		
A (4.0 Pts) Excellent	F (0.0 Pts) Failing	A (4.0 Pts) Excellent	F (0.0 Pts)	Failing
B (3.0 Pts) Good	P (4.0 Pts) Passed (Pass/Fail Option)	B (3.0 Pts) Good	P (4.0 Pts)	Passed (Pass/Fail Option)
C (0.0 Pts) Unsatisfactory	WF (0.0 Pts) Withdrawn - Failing	C (2.0 Pts) Satisfactory	WF (0.0 Pts)	Withdrawn - Failing
D (0.0 Pts) Unsatisfactory		D (0 Pts) Unsatisfactory		

Repeated Courses

Repeated courses are counted in the John Patrick University of Health and Applied Sciences grade point average and may also be counted in the student's primary program GPA (Student Program GPA), depending on the policies of the student's program. The first attempt to complete a course is listed as attempted credits not earned. The following grades are not considered in computing semester or cumulative grade point averages:

- AU Audit No Credit
- I Incomplete/Pending
- T Denotes credits transferred from another Institution
- W Withdrawn R Repeated Course

Abbreviations and Symbols

- EHRS Credit hours earned
- OPts Quality Points Earned
- GPA Grade point average (computed by dividing QPts by EHRS)

Credit Types

Regular Credit - All John Patrick University of Health and Applied Sciences credit is reported in terms of semester hours.

II. Record Format

The "Official Transcript" standard format lists course history, grade and GPA information in chronological order sorted by the student's career level. The "Official Transcript with Enrollment" provides the same information as the standard transcript but also includes all courses in which a student is currently enrolled or registered. "Official Transcript" or "Official Transcript with Enrollment" (Without career level designation) indicates that the document contains all work completed at John Patrick University of Health and Applied Sciences.

The JPU GPA reflects the student's GPA according to standard university-wide rules. A Semester JPU GPA and a cumulative to date JPU GPA are calculated at the end of each semester. The overall JPU GPA summary statistics are reflected at the end of each student career level.

The Student Program GPA is calculated according to the rules determined by the student's primary academic program at the time of printing. The cumulative Student Program GPA summary statistics are reflected at the end of each student career level and are based on the student's last active primary program at that level.

III. Transfer, Test and Special Credit

Courses accepted in transfer from other institutions are listed under a Transfer Credit heading. Generally, a grade of "T" (transfer grade) is assigned and course numbers, titles and credit hours assigned reflect JPU Equivalents. Transfer hours with a grade of "T" are not reflected in the cumulative grade averages; however, the hours are included in the "Hrs Earned" Field.

IV.Accreditation

This Institution is authorized by: the Indiana Board for Proprietary Education, 101 West Ohio Street, Suite 300 Indianapolis, Indiana 46204-4206. Phone (317) 464-4400 Ext. 138.

This Institution is accredited by the Accrediting Commission of Career Schools and Colleges (ACCSC), 2101 Wilson Boulevard, Suite 302 Arlington, VA 22201. Phone (703) 247-4212. Website: www.accsc.org. ACCSC is recognized by the United States Department of Education.

This Institution holds programmatic accreditation by the Joint Review Committee on Education in Radiologic Technology (JRCERT), 20 North Wacker Drive, Suite 2850 Chicago, Illinois 60606-3182. Phone (312) 704-5300. Email: mail@jrcert.org. Programs Accredited: Bachelor of Science in Medical Dosimetry and Master of Science in Medical Dosimetry.

V. Validation

A transcript issued by John Patrick University of Health and Applied Sciences is official when it displays signatures. Printed official transcripts display signatures and are printed on SCRIP-SAFE Security paper. A raised seal is not required.

VI. Registrar Contact

Questions about the content of this record should be referred to the Office of the Registrar where it was printed.