



Radiation Physics
CONSULTANTS

July 18, 2023

MQSA PHYSICIST CONTINUING EXPERIENCE

This document expires on May 8, 2025.

This letter is in reference to the Continuing Experience requirement set forth by the MQSA regarding the number of at least 6 mammography surveys completed at least 2 sites within a 24-month period.

Steven T. Nicholas, M.S., a Radiation Physics Consultants, Inc. physicist, has met the Continuing Experience requirements of MQSA. Surveys at the following facilities have been completed within the required time frame:

Site	City	Unit	Date
St. Luke's Hospital	Duluth, MN	Selenia Dimensions DBT	May 8, 2023
Lakewood Health Center	Baudette, MN	Siemens Revelations DBT	May 24, 2023
DMS Health Technologies	Mobile Truck	Selenia Dimensions DBT	June 2, 2023
Mariner Medical Clinic	Superior, WI	Selenia Dimensions DBT	June 27, 2023
Osceola Medical Center	Osceola, WI	Selenia Dimensions DBT	June 13, 2023
North Shore Health	Grand Marais, MN	Selenia Dimensions DBT	July 6, 2023

Please do not hesitate to contact me if you have any additional questions.


Sincerely,

Steven T. Nicholas, M.S., DABMP
President, RPC



MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

Full-Field Digital – Lorad/Hologic

Site Name	North Shore Health	Report Date	7/6/2023
Address	515 5th Ave West, Grand Marais, MN 55604	Survey Date	7/6/2023
Medical Physicist's Name	Steven T. Nicholas	Signature	
X-Ray Unit Manufacturer	Lorad/Hologic	Model	Selenia Dimensions DBT
Date of Installation	6/21/2019	Room ID	Mammo
		SN	SDM131900754
QC Manual Version #	MAN-03706, Rev. 010 (Aug 2020) <small>(use any version applicable to model; contact mfr if questions)</small>		

Accessory Equipment	Manufacturer	Model	Location	QC Manual Version #
Review Workstation*	Barco/Hologic	MDMG-5221	<input type="checkbox"/> On-site <input type="checkbox"/> Off-site	MAN-04959, Rev. 002
Film Printer*	NA	NA	NA	NA

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polc_Guidance_Help_System.htm).

Survey Type- MEE following AEC recalibration and Annual Survey
Features- 2D and Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

		PASS/FAIL
1. Mammographic Unit Assembly Evaluation		Pass
2. Collimation Assessment		Pass
3. Artifact Evaluation		Pass
4. kVp Accuracy and Reproducibility		Pass
5. Beam Quality Assessment - HVL Measurement		Pass
6. Evaluation of System Resolution		Pass
7. Automatic Exposure Control (AEC) Function Performance <small>(NA for systems without AEC)</small>		Pass
8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose		
Average glandular dose for average breast is ≤ 3 mGy (300 mrad) <small>(conventional)</small>	118	mrad
Average glandular dose for average breast is ≤ 3 mGy (300 mrad) (DBT)	148	mrad
9. Radiation Output Rate		Pass
10. Phantom Image Quality Evaluation		
Phantom image scores <small>(conventional)</small>	6.0	4.0
Phantom image scores (DBT)	6.0	4.0
	4.5	4.5
11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements <small>(values required for all tests)</small>		
SNR <small>(value)</small>	56.7	
CNR <small>(value)</small>	11.16	<small>(required for new unit MEE and Annual Survey)</small>
CNR should not vary by more than $\pm 15\%$ <small>(NA for MEE)</small>		
12. Diagnostic Review Workstation (RWS) QC <small>(for all RWS, even if located offsite; NA if only hardcopy read)</small>		Pass
13. DICOM Printer QC <small>(if applicable, MEE only)</small>		NA
14. Detector Flat Field Calibration <small>(MEE only)</small>		Pass
15. Geometry Calibration For Tomosynthesis (DBT MEE only)		Pass
16. Compression Thickness Indicator <small>(MEE only)</small>		Pass
17. Compression <small>(MEE only)</small>		Pass
18. Detector Ghosting <small>(troubleshooting only)</small>		NA

***** YOUR MEDICAL PHYSICIST MUST SUMMARIZE HIS/HER RESULTS ON THIS FORM *****

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

(Lorad, continued)

Evaluation of Site's Technologist QC Program

	Frequency	PASS/FAIL
1. DICOM Printer Quality Control <i>(if applicable)</i>	Weekly	NA
2. Viewboxes and Viewing Conditions	Weekly	Pass
3. Artifact Evaluation	Weekly	Pass
4. Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	Pass
5. Phantom Image Quality Evaluation	Weekly	Pass
6. Detector Flat-Field Calibration	Weekly	Pass
7. Compression Thickness Indicator	Bi-weekly	Pass
8. Visual Checklist	Monthly	Pass
9. Repeat/Reject Analysis	Quarterly	Pass
10. Compression	Semi-annually	Pass
11. Geometry Calibration (Tomosynthesis Option) (DBT)	Semi-annually	Pass
12. Diagnostic Review Workstation QC <i>(NA if only hardcopy read)</i>	See Hologic QC Manual	Pass
13. Mobile Unit Quality Control <i>(if applicable)</i>	After every move	NA

Medical Physicist's Recommendations for Quality Improvement

This is an annual Medical Physicist's survey on a DBT unit.

Medical Physicist's QC Tests
No Discrepancies.


Evaluation of Site's Technologist QC Program
There are no discrepancies.

Site does not print.

Important: The FDA's MQSA regulations require that the facility's quality assurance program for digital mammography be "substantially the same as the quality assurance program recommended by the image receptor manufacturer." The medical physicist should follow the QC manual version provided by the manufacturer for the unit surveyed. See the attached FDA Approved Alternative Requirement for the Lorad Selenia for the allowed correction periods on taking corrective action for test failures. (If these tests are performed as part of an Equipment Evaluation, corrective action must be taken before mammographic images are acquired.)

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

Full-Field Digital – Lorad/Hologic

Site Name	Mariner Medical Clinic	Report Date	7/18/2023
Address	109 North 28th St, Superior, WI 54880	Survey Date	6/27/2023
Medical Physicist's Name	Steven T. Nicholas	Signature	
X-Ray Unit Manufacturer	Lorad/Hologic	Model	Selenia Dimensions DBT
Date of Installation	9/11/2020	Room ID	Mammo
		SN	SDM131901270
QC Manual Version #	MAN-03706, Rev. 009 (Sept. 2019) <small>(use any version applicable to model; contact mfr if questions)</small>		

Accessory Equipment	Manufacturer	Model	Location	QC Manual Version #
Review Workstation*	Barco/Hologic	MDMG-5221	<input type="checkbox"/> On-site <input type="checkbox"/> Off-site	MAN-04959, Rev. 002
Film Printer*	NA	NA	NA	NA

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polig_Guidance_Help_System.htm).

Survey Type- MEE for new detector
Features- 2D and Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

		PASS/FAIL
1. Mammographic Unit Assembly Evaluation		Pass
2. Collimation Assessment		Pass
3. Artifact Evaluation		Pass
4. kVp Accuracy and Reproducibility		Pass
5. Beam Quality Assessment - HVL Measurement		Pass
6. Evaluation of System Resolution		Pass
7. Automatic Exposure Control (AEC) Function Performance <small>(NA for systems without AEC)</small>		Pass
8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose		
Average glandular dose for average breast is ≤ 3 mGy (300 mrad) <small>(conventional)</small>	133	mrad
Average glandular dose for average breast is ≤ 3 mGy (300 mrad) (DBT)	156	mrad
9. Radiation Output Rate		Pass
10. Phantom Image Quality Evaluation		
Phantom image scores <small>(conventional)</small>	6.0	4.0
Phantom image scores (DBT)	6.0	4.0
11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements <small>(values required for all tests)</small>		
SNR <small>(value)</small>	57.8	
CNR <small>(value)</small>	11.35	<small>(required for new unit MEE and Annual Survey)</small>
CNR should not vary by more than $\pm 15\%$ <small>(NA for MEE)</small>		
12. Diagnostic Review Workstation (RWS) QC <small>(for all RWS, even if located offsite; NA if only hardcopy read)</small>		Pass
13. DICOM Printer QC <small>(if applicable, MEE only)</small>		NA
14. Detector Flat Field Calibration <small>(MEE only)</small>		Pass
15. Geometry Calibration For Tomosynthesis (DBT MEE only)		Pass
16. Compression Thickness Indicator <small>(MEE only)</small>		Pass
17. Compression <small>(MEE only)</small>		Pass
18. Detector Ghosting <small>(troubleshooting only)</small>		NA


***** YOUR MEDICAL PHYSICIST MUST SUMMARIZE HIS/HER RESULTS ON THIS FORM *****

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

(Lorad, continued)

Evaluation of Site's Technologist QC Program

(Required for Annual Surveys; not required for Mammography Equipment Evaluations of new units. However, medical physicists **must** review the site's technologist QC program within 45 days and complete this section so that the facility may submit this form along with the entire Mammography Equipment Evaluation report with their phantom and clinical images to the ACR.)

	Frequency	PASS/FAIL
1. DICOM Printer Quality Control (if applicable)	Weekly	NA
2. Viewboxes and Viewing Conditions	Weekly	NA
3. Artifact Evaluation 	Weekly	NA
4. Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	NA
5. Phantom Image Quality Evaluation	Weekly	NA
6. Detector Flat-Field Calibration	Weekly	NA
7. Compression Thickness Indicator	Bi-weekly	NA
8. Visual Checklist	Monthly	NA
9. Repeat/Reject Analysis	Quarterly	NA
10. Compression	Semi-annually	NA
11. Geometry Calibration (Tomosynthesis Option) (DBT)	Semi-annually	NA
12. Diagnostic Review Workstation QC (NA if only hardcopy read)	See Hologic QC Manual	NA
13. Mobile Unit Quality Control (if applicable)	After every move	NA

Medical Physicist's Recommendations for Quality Improvement

This is Medical Physicist's MEE following replacement of the detector array.

Medical Physicist's QC Tests

No Discrepancies.


Evaluation of Site's Technologist QC Program

Not performed. Keep your current baselines.

Important: The FDA's MQSA regulations require that the facility's quality assurance program for digital mammography be "substantially the same as the quality assurance program recommended by the image receptor manufacturer." The medical physicist should follow the QC manual version provided by the manufacturer for the unit surveyed. See the attached FDA Approved Alternative Requirement for the Lorad Selenia for the allowed correction periods on taking corrective action for test failures. (If these tests are performed as part of an Equipment Evaluation, corrective action must be taken before mammographic images are acquired.)

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

Full-Field Digital – Lorad/Hologic

Site Name	Osceola Medical Center	Report Date	6/14/2023
Address	2600 65th Ave, Osceola, WI 54020	Survey Date	6/13/2023
Medical Physicist's Name	Steven T. Nicholas	Signature	
X-Ray Unit Manufacturer	Lorad/Hologic	Model	Selenia Dimensions DBT
Date of Installation	9/20/2018	Room ID	Mammo Rm 1
		SN	SDM131500755
QC Manual Version #	MAN-03706, Rev. 009 (Sept 2019) <small>(use any version applicable to model; contact mfr if questions)</small>		

Accessory Equipment	Manufacturer	Model	Location	QC Manual Version #
Review Workstation*	Barco	MDMG-5221	On-Site	MAN-03706, Rev. 009 (Sept 2019)
Film Printer*	NA	NA	NA	NA

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polc_Guidance_Help_System.htm).

Survey Type- Annual Survey
Features- 2D and Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)


	PASS/FAIL
1. Mammographic Unit Assembly Evaluation	Pass
2. Collimation Assessment	Pass
3. Artifact Evaluation	Pass
4. kVp Accuracy and Reproducibility	Pass
5. Beam Quality Assessment - HVL Measurement	Pass
6. Evaluation of System Resolution	Pass
7. Automatic Exposure Control (AEC) Function Performance <small>(NA for systems without AEC)</small>	Pass
8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose	
Average glandular dose for average breast is ≤ 3 mGy (300 mrad) <small>(conventional)</small>	124 mrad
Average glandular dose for average breast is ≤ 3 mGy (300 mrad) (DBT)	144 mrad
9. Radiation Output Rate	Pass
10. Phantom Image Quality Evaluation	
Phantom image scores <small>(conventional)</small>	Pass
Phantom image scores (DBT)	Pass
11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements <small>(values required for all tests)</small>	
SNR <small>(value)</small>	56.9
CNR <small>(value)</small>	10.73
<small>(required for new unit MEE and Annual Survey)</small>	
CNR should not vary by more than $\pm 15\%$ <small>(NA for MEE)</small>	Pass
12. Diagnostic Review Workstation (RWS) QC <small>(for all RWS, even if located offsite; NA if only hardcopy read)</small>	Pass
13. DICOM Printer QC <small>(if applicable, MEE only)</small>	NA
14. Detector Flat Field Calibration <small>(MEE only)</small>	Pass
15. Geometry Calibration For Tomosynthesis (DBT MEE only)	Pass
16. Compression Thickness Indicator <small>(MEE only)</small>	Pass
17. Compression <small>(MEE only)</small>	Pass
18. Detector Ghosting <small>(troubleshooting only)</small>	NA

***** YOUR MEDICAL PHYSICIST MUST SUMMARIZE HIS/HER RESULTS ON THIS FORM *****

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

(Lorad, continued)

Evaluation of Site's Technologist QC Program

	Frequency	PASS/FAIL
1. DICOM Printer Quality Control <i>(if applicable)</i>	Weekly	NA
2. Viewboxes and Viewing Conditions	Weekly	Pass
3. Artifact Evaluation	Weekly	Pass
4. Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	Pass
5. Phantom Image Quality Evaluation	Weekly	Pass
6. Detector Flat-Field Calibration	Weekly	Pass
7. Compression Thickness Indicator	Bi-weekly	Pass
8. Visual Checklist	Monthly	Pass
9. Repeat/Reject Analysis	Quarterly	Pass
10. Compression	Semi-annually	Pass
11. Geometry Calibration (Tomosynthesis Option) (DBT) 	Semi-annually	Pass
12. Diagnostic Review Workstation QC <i>(NA if only hardcopy read)</i>	See Hologic QC Manual	Pass
13. Mobile Unit Quality Control <i>(if applicable)</i>	After every move	NA


Medical Physicist's Recommendations for Quality Improvement

<p>This is an annual Medical Physicist's survey.</p> <p>Medical Physicist's QC Tests</p> <p>No Discrepancies.</p> <p>Evaluation of Site's Technologist QC Program</p> <p>No discrepancies.</p> <p>Site does not print.</p>
--

<p>Important: The FDA's MQSA regulations require that the facility's quality assurance program for digital mammography be "substantially the same as the quality assurance program recommended by the image receptor manufacturer." The medical physicist should follow the QC manual version provided by the manufacturer for the unit surveyed. See the attached FDA Approved Alternative Requirement for the Lorad Selenia for the allowed correction periods on taking corrective action for test failures. (If these tests are performed as part of an Equipment Evaluation, corrective action must be taken before mammographic images are acquired.)</p>
--

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

Full-Field Digital – Lorad/Hologic

Site Name	DMS Health Technologies	Report Date	6/7/2023
Address	728 Benton Drive, Suite 101, West Fargo, ND 58078	Survey Date	6/2/2023
Medical Physicist's Name	Steven T. Nicholas	Signature	
X-Ray Unit Manufacturer	Lorad/Hologic	Model	Selenia Dimensions DBT
Date of Installation	5/24/2022	Room ID	MM28
		SN	SDM131901781
QC Manual Version #	MAN-03706, Rev. 011 <small>(use any version applicable to model; contact mfr if questions)</small>		

Accessory Equipment	Manufacturer	Model	Location	QC Manual Version #
Review Workstation*	Hologic	SecurView	Off-Site	MAN-03706, Rev. 011
Film Printer*	NA	NA	NA	NA

**FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polig_Guidance_Help_System.htm).*

Survey Type- MEE following an upgrade and AEC recalibration. A full annual was also performed.
Features- 2D and Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)


		PASS/FAIL
1. Mammographic Unit Assembly Evaluation		Pass
2. Collimation Assessment		Pass
3. Artifact Evaluation		Pass
4. kVp Accuracy and Reproducibility		Pass
5. Beam Quality Assessment - HVL Measurement		Pass
6. Evaluation of System Resolution		Pass
7. Automatic Exposure Control (AEC) Function Performance <i>(NA for systems without AEC)</i>		Pass
8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose		
Average glandular dose for average breast is ≤ 3 mGy (300 mrad) <i>(conventional)</i>	31	mrad
Average glandular dose for average breast is ≤ 3 mGy (300 mrad) (DBT)	52	mrad
9. Radiation Output Rate		Pass
10. Phantom Image Quality Evaluation		
Phantom image scores <i>(conventional)</i>	5.5	5.0
Phantom image scores (DBT)	5.5	4.0
	4.5	4.5
11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements <i>(values required for all tests)</i>		
SNR <i>(value)</i>	59.2	
CNR <i>(value)</i>	10.70	<i>(required for new unit MEE and Annual Survey)</i>
CNR should not vary by more than $\pm 15\%$ <i>(NA for MEE)</i>		
12. Diagnostic Review Workstation (RWS) QC <i>(for all RWS, even if located offsite; NA if only hardcopy read)</i>		Pass
13. DICOM Printer QC <i>(if applicable, MEE only)</i>		NA
14. Detector Flat Field Calibration <i>(MEE only)</i>		NA
15. Geometry Calibration For Tomosynthesis (DBT MEE only)		NA
16. Compression Thickness Indicator <i>(MEE only)</i>		Pass
17. Compression <i>(MEE only)</i>		NA
18. Detector Ghosting <i>(troubleshooting only)</i>		NA

***** YOUR MEDICAL PHYSICIST MUST SUMMARIZE HIS/HER RESULTS ON THIS FORM *****

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

(Lorad, continued)

Evaluation of Site's Technologist QC Program

	Frequency	PASS/FAIL
1. DICOM Printer Quality Control <i>(if applicable)</i>	Weekly	NA
2. Viewboxes and Viewing Conditions	Weekly	Pass
3. Artifact Evaluation	Weekly	Pass
4. Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	Pass
5. Phantom Image Quality Evaluation	Weekly	Pass
6. Detector Flat-Field Calibration	Weekly	Pass
7. Compression Thickness Indicator	Bi-weekly	Pass
8. Visual Checklist	Monthly	Pass
9. Repeat/Reject Analysis	Quarterly	Pass
10. Compression	Semi-annually	Pass
11. Geometry Calibration (Tomosynthesis Option) (DBT) 	Semi-annually	Pass
12. Diagnostic Review Workstation QC <i>(NA if only hardcopy read)</i>	See Hologic QC Manual	Pass
13. Mobile Unit Quality Control <i>(if applicable)</i>	After every move	Pass

Medical Physicist's Recommendations for Quality Improvement

This is an Annual survey of a mobile DBT unit.

Medical Physicist's QC Tests
No Discrepancies.

Evaluation of Site's Technologist QC Program
No discrepancies.

Important: The FDA's MQSA regulations require that the facility's quality assurance program for digital mammography be "substantially the same as the quality assurance program recommended by the image receptor manufacturer." The medical physicist should follow the QC manual version provided by the manufacturer for the unit surveyed. See the attached FDA Approved Alternative Requirement for the Lorad Selenia for the allowed correction periods on taking corrective action for test failures. (If these tests are performed as part of an Equipment Evaluation, corrective action must be taken before mammographic images are acquired.)

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

Full-Field Digital – Lorad

Site Name	St. Luke's Hospital	Report Date	7/18/2023
Address	915 E 1st St, Duluth, MN 55805	Survey Date	5/8/2023
Medical Physicist's Name	Steven T. Nicholas	Signature	
X-Ray Unit Manufacturer	Hologic	Model	Selenia Dimensions DBT
Date of Installation	3/10/2017	Room ID	Mammo Room 1
		SN	81002177895
QC Manual Version #	MAN-03706, Rev. 006 (June 2017) <small>(use any version applicable to model; contact mfr if questions)</small>		

Accessory Equipment	Manufacturer	Model	Location	QC Manual Version #
Review Workstation*	Barco/Hologic	MDMG-5221	On-site	MAN-04959, Rev. 002
Film Printer*	NA	NA	NA	NA

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System www.fda.gov/CDRH/MAMMOGRAPHY/robohelp/START.HTM.

Survey Type Annual Survey of 2D and Tomo

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)


	PASS/FAIL
1. Mammographic Unit Assembly Evaluation	Pass
2. Collimation Assessment	
Deviation between X-ray field and light field ≤ 2% of SID	Pass
X-ray field does not extend beyond any side of the IR by >2% of SID	Pass
X-ray field covers all of the IR on the chest wall side	Pass
Paddle chest wall edge does not extend beyond IR by >1% of SID or appear on mammogram	Pass
3. Artifact Evaluation <small>(no significant artifacts visible)</small>	Pass
4. kVp Accuracy and Reproducibility	
Measured average kVp within ±5% of indicated kVp	Pass
kVp coefficient of variation ≤ 0.02	Pass
5. Beam Quality Assessment - HVL Measurement	Pass
6. Evaluation of System Resolution <small>(system limiting spatial resolution >7 cycles/mm (lp/mm))</small>	Pass
7. Automatic Exposure Control (AEC) Function Performance <small>(NA for systems without AEC)</small>	
Each pixel value (2-8 cm; all operating modes) within ±10% of mean	Pass
Exposure compensation steps performance within acceptable limits	Pass
8. Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose	
Average glandular dose for average breast is ≤3 mGy (300 mrad)	Pass
Average glandular to a 4.2 cm breast on your unit is	
mrad 117 Conventional	
mrad 132 Tomosynthesis Option	
Coefficient of variation for either R or mAs ≤ 0.05 (NA for systems without AEC)	Pass
9. Radiation Output Rate <small>(> 230 mR/sec)</small>	
mR/sec 700	Pass
10. Phantom Image Quality Evaluation	
Phantom image quality scores (Conventional)	
Fibers 6.0 Specks 4.0 Masses 4.5 Pass	
Phantom image quality scores (Tomosynthesis Option)	
Fibers 6.0 Specks 4.0 Masses 4.5 Pass	
11. Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements <small>(values required for all tests)</small>	
SNR (value) 52.0	Pass
CNR (value) 10.5 <small>(Required for both new unit Mammography Equipment Evaluations and Annual Surveys)</small>	
CNR should not vary by more than ±15% (NA for Equipment Evaluation)	Pass
12. Diagnostic Review Workstation (RWS) QC <small>(for all RWS, even if located offsite; NA if only hardcopy read)</small>	Pass
13. DICOM Printer QC <small>(Mammography Equipment Evaluations only)</small>	NA
14. Detector Flat Field Calibration <small>(Mammography Equipment Evaluations only)</small>	NA
15. Compression Thickness Indicator <small>(Mammography Equipment Evaluations only)</small>	Pass
16. Compression <small>(Mammography Equipment Evaluations only)</small>	NA
17. Geometry Calibration	NA

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

(Lorad, continued)

Evaluation of Site's Technologist QC Program

(Required for Annual Surveys; not required for Mammography Equipment Evaluations of new units. However, medical physicists **must** review the site's technologist QC program within 45 days and complete this section so that the facility may submit this form along with the entire Mammography Equipment Evaluation report with their phantom and clinical images to the ACR.)

	Frequency	PASS/FAIL
1. DICOM Laser Printer Quality Control	Weekly	NA
2. Detector Flat-Field Calibration	Weekly	Pass
3. Geometry Calibration (Tomosynthesis Option) 	Weekly	Pass
4. Artifact Evaluation	Weekly	Pass
5. Phantom Image Quality Evaluation	Weekly	Pass
6. Signal-To-Noise and Contrast-To-Noise Measurements	Weekly	Pass
7. Compression Thickness Indicator	Bi-weekly	Pass
8. Review Workstation QC-Overall	See FDA guidance	Pass
9. Viewboxes and Viewing Conditions	Weekly	Pass
10. Visual Checklist	Monthly	Pass
11. Repeat Analysis	Quarterly	Pass
12. Compression	Semi-annually	Pass

Medical Physicist's Recommendations for Quality Improvement

This is an annual survey.

Medical Physicist's QC Tests
No Discrepancies.

Evaluation of Site's Technologist QC Program
No discrepancies. (QC for both rooms was reviewed at this time. Will likely review both again in June. I intend to move everything to June for 2024).

This facility does not print hard copy.

Important: The FDA's MQSA regulations require that the facility's quality assurance program for digital mammography be "substantially the same as the quality assurance program recommended by the image receptor manufacturer." The medical physicist should follow the QC manual version provided by the manufacturer for the unit surveyed. See the attached FDA Approved Alternative Requirement for the Lorad Selenia for the allowed correction periods on taking corrective action for test failures. (If these tests are performed as part of an Equipment Evaluation, corrective action must be taken before mammographic images are acquired.)

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

Full-Field Digital – Siemens

Site Name	LakeWood Health Center	Report Date	5/31/2023
Address	600 Main Ave S., Baudette, MN 56623	Survey Date	5/24/2023
Medical Physicist's Name	Steven T. Nicholas	Signature	
X-Ray Unit Manufacturer	Siemens	Model	Mammomat Revelation Tomosynthesis
Date of Installation	6/10/2021	Room ID	DBT Mammo

QC Manual Version # Tomo QC VC20, 2D QC VC20 (use version applicable to unit tested; contact mfr if questions)

Accessory Equipment	Manufacturer	Model	Location	QC Manual Version #
Review Workstation*	Double Black	Wide 5MP	<input type="checkbox"/> On-site <input type="checkbox"/> Off-site	X-Cal Software 5.2.0.5
Film Printer*	NA	NA	NA	NA

*FDA recommends that only monitors and printers specifically cleared for FFDM use by FDA's Office of Device Evaluation (ODE) be used. See FDA's Policy Guidance Help System (www.accessdata.fda.gov/cdrh_docs/presentations/pghs/Polic_Guidance_Help_System.htm).

Survey Type Mammo Eqpt Evaluation (MEE) of new unit (include MQSA Rqmts for Mammo Eqpt checklist) Annual Survey
Features 2D Digital Breast Tomosynthesis (DBT)

Medical Physicist's QC Tests

("Pass" means all components of the test passes; indicate "Fail" if any component fails. Tests must be done for both on and off-site equipment.)

	PASS/FAIL
1. Image Quality	Pass
<small>Largest 5 fibers, 4 speck groups and 4 masses visible*</small>	
<small>(*largest 4 fibers, 3 speck groups and 3 masses acceptable if spatial resolution and CNR pass)</small>	
Phantom image scores: Fibers 6.0 Specks 4.0 Masses 4.5	
2. Artifact Detection	Pass
3. Printer Check <small>(if applicable)</small>	NA
4. SNR, CNR and AEC Repeatability	Pass
Measured values: SNR 63.1 CNR 2.65	
CV for mAs and entrance air kerma $\leq 5\%$	Pass
Max deviation of mean pixel values and SNR within $\pm 15\%$ of mean for measurements	Pass
5. Radiation Dose	Pass
Average glandular dose for average breast is ≤ 3 mGy (300 mrad)	
2D 0.80 mGy	
3D 1.70 mGy	
2D + 3D 2.50 mGy	
6. Spatial Resolution	Pass
7. AEC Test	Pass
8. Detector Uniformity	Pass
9. Mechanical Tests	Pass
10. Acquisition Workstation Monitor Check	Pass
11. Site Audit/Evaluation of Technologist QC Program	Pass
12. Collimation, Dead Space & Compression Paddle Position	Pass
13. HVL and Radiation Output	Pass
14. Tube Voltage Measurement & Repeatability	Pass
15. Average Glandular Dose (DBT)	Pass
3D 1.70 mGy	
16. Geometric Accuracy in X and Y Direction and Z-Resolution (DBT)	Pass
17. Radiation Field (DBT)	Pass
18. System Imaging Quality (DBT)	Pass
<small>≥ 4 fibers, ≥ 3 speck groups and ≥ 3 masses must be visible</small>	
Phantom image scores: Fibers 6.0 Specks 4.0 Masses 4.0	
19. Artifact Detection (DBT)	Pass
20. Review Workstation (RWS) Tests <small>(for all RWS, even if located offsite; NA if only hardcopy read)</small>	Pass

***** YOUR MEDICAL PHYSICIST MUST SUMMARIZE HIS/HER RESULTS ON THIS FORM *****

MEDICAL PHYSICIST'S MAMMOGRAPHY QC TEST SUMMARY

(Siemens, continued)

Evaluation of Technologist QC Program

New units: Medical physicists **must** review the technologist QC **within 45 days of installation** and complete this section. The facility is required to submit the entire Mammography Equipment Evaluation report (including this form) along with their testing materials for accreditation.

Existing units: Medical physicists **must** complete this section as part of the unit's annual survey.

Relocating units: This section is **not** required if the medical physicist does **not** conduct a complete annual survey after relocation.

		FREQUENCY	PASS/FAIL
1.	Phantom Image Quality	Novation & Fusion-Daily; Inspiration-Weekly	Pass
2.	Artifact Detection	Weekly	Pass
3.	SNR and CNR Measurements	Weekly	Pass
4.	Detector Calibration	Novation-Weekly; Inspiration & Fusion-Quarterly	Pass
5.	Repeat/Reject Analysis	Quarterly	Pass
6.	Compression Force	Semi-annually	Pass
7.	System Imaging Quality (DBT)	Daily when DBT performed	Pass
8.	Printer Check (if applicable)	Daily, when images printed	NA
9.	Review Workstation QC-Overall (NA if only hardcopy read)	See FDA guidance	Pass
10.	Mobile Unit Quality Control (if applicable)	After every move	NA

Medical Physicist's Recommendations for Quality Improvement

This is an annual survey. There are no discrepancies with the machine or QC program.

Important:

1. The facility's "quality assurance program shall be ***substantially the same*** as the quality assurance program recommended by the ***image receptor [digital detector] manufacturer***." This is required by the FDA.
2. Use the QC manual version provided by the manufacturer ***for the digital system surveyed***.
3. If the RWS or printer is FDA-cleared for FFDM, their ***QC manual*** is considered to be "***substantially the same***" and may be followed. (Check with the RWS or printer manufacturers for their clearance status and QC manual.)
4. If the RWS or printer is not cleared by the FDA for FFDM, ***follow the QC manual provided by the image receptor manufacturer***. (Check with the image receptor manufacturer for their required tests.)
5. All tests must be evaluated for the facility's ***on and off-site*** equipment. If the evaluation was done on a different day than the survey date, note the date above.
6. See the FDA-approved alternative standard for Siemens FFDM regarding corrective action periods when components fail QC. However, if these tests are performed as part of a Mammography Equipment Evaluation (e.g., for a new system), corrective action must be taken before mammographic images are acquired.